



Certificate of Traceability

Manufacturer Certification

This certifies that This certifies that MedTestDx, using Pointe Scientific, HbA1c Immunoturbidimetric on BS480 has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: European Reference Laboratory ESRL#13

The system evaluated was:

Instrument: BS480	Calibrator Lot: 906004
Reagent Lot: 834701	Calibrator Assigned Values: 0.00%, 4.8%, 7.3%, 11.3%, 15.2%

Date of Certification: July 1, 2019

Certification Expires: July 1, 2020

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor