

1. IDENTIFICATION AND PRODUCT INFORMATION

Product Use: Urinalysis Drugs of Abuse (DAU) Calibrators and Controls

Supplier: MedTest Dx
 5449 Research Drive
 Canton, Michigan 48188
 Ph: 1-734-487-8300
 Fax: 1-734-483-1592
 Email: techservice@medtestdx.com

Emergency: MedTest Dx Tel. 1-800-757-5313

Reference Number:	60-UN0002 60-AC0598, 60-AC0584, 60-AC0586, 60-AC0588, 60-AC0590, 60-AC0594, 60-AC0596 60-AM0204, 60-AM0206, 60-AM0208, 60-AM0210, 60-AM0214, 60-AM0216 60-AM0084, 60-AM0086, 60-AM0088, 60-AM0090, 60-AM0094, 60-AM0096 60-BA0284, 60-BA0286, 60-BA0288, 60-BA0290, 60-BA0292 60-BZ0264, 60-BZ0266, 60-BZ0268, 60-BZ0270, 60-BZ0272 60-NB0544, 60-NB0546, 60-NB0548, 60-NB0550, 60-NB0554, 60-NB0556, 60-NB0558 60-NB0552 60-MB0744, 60-MB0746, 60-MB0748, 60-MB0750, 60-MB0754, 60-MB0756 60-CC0684, 60-CC0686, 60-CC0688, 60-CC0690, 60-CC0694, 60-CC0696 60-CC0064, 60-CC0066, 60-CC0068, 60-CC0070, 60-CC0074, 60-CC0076 60-EC0324, 60-EC0326, 60-EC0328, 60-EC0330, 60-EC0334, 60-EC0336 60-EA0446, 60-EA0448, 60-EA-0450 60-NF0626, 60-NF0634, 60-NF0636 60-HY0764, 60-HY0766, 60-HY0768, 60-HY0770, 60-HY0774, 60-HY0776 60-HY0784, 60-HY0786, 60-HY0788, 60-HY0790, 60-HY0794, 60-HY0796 60-ME0224, 60-ME0226, 60-ME0228, 60-ME0230, 60-ME0234, 60-ME0236 60-MM0404, 60-MM0406, 60-MM0408, 60-MM0410, 60-MM0414, 60-MM0416 60-MM0384, 60-MM0386, 60-MM0388, 60-MM0390, 60-MM0394, 60-MM0396 60-MP0704, 60-MP0706, 60-MP0708, 60-MP0710, 60-MP0714, 60-MP0716 60-MA1704, 60-MA1706, 60-MA1708, 60-MA1710, 60-MA1714, 60-MA1716 60-MA1742, 60-MA1744, 60-MA1746, 60-MA1748, 60-MA1750, 60-MA1754, 60-MA1756 60-MA1782, 60-MA1784, 60-MA1786, 60-MA1788, 60-MA1790, 60-MA1794, 60-MA1796 60-OP0044, 60-OP0046, 60-OP0048, 60-OP0050, 60-OP0054, 60-OP0056 60-OP0664, 60-OP0666, 60-OP0668, 60-OP0670, 60-OP0674, 60-OP0676 60-OX0504, 60-OX0484, 60-OX0486, 60-OX0488, 60-OX0492, 60-OX0496, 60-OX0498 60-OX0490, 60-OX0494 60-PH0024, 60-PH0026, 60-PH0028, 60-PH0030, 60-PH0034, 60-PH0036 60-PP0244, 60-PP0246, 60-PP0248, 60-PP0250, 60-PP0254, 60-PP0256 60-TH0004, 60-TH0144, 60-TH0146, 60-TH0014, 60-TH0150, 60-TH0012, 60-TH0148 60-TH0152, 60-TH0154, 60-TH0016, 60-TH0156, 60-TH0158, 60-TH0018	Universal Negative Calibrator 6-Acetylmorphine Amphetamines 500 Amphetamines Barbiturate Benzodiazepine Norbuprenorphine Carisoprodol Metabolite (Meprobamate) Cocaine Metabolite 150 Cocaine Metabolite Ecstasy (MDMA) Ethyl Alcohol Norfentanyl (Q) Hydrocodone 100 Hydrocodone 300 Methadone Methadone Metabolite (EDDP) 100 Methadone Metabolite (EDDP) 300 Methamphetamine Multi-Analyte Set C Multi-Analyte Set D Multi-Analyte Set E Opiate Opiate 2000 Oxycodone Phencyclidine Propoxyphene THC (Cannabinoid)
--------------------------	---	--

Description: Urine based calibrators and controls for use in combination with drugs of abuse (DAU) urinalysis enzyme immunoassay (EIA) and enzymatic assay diagnostic reagents. The products are contained in dropper bottles and packaged in small boxes. For further information, please go to the product insert included inside of the assay box or go to www.medtestdx.com for the product inserts specifically for calibrators and/or controls.

2. HAZARD(S) IDENTIFICATION

Classification:	The product is not considered a hazardous mixture, but it is potentially biohazardous because it contains processed human urine. It contains a trace quantity amount (< 0.1 % w/v) of sodium azide (NaN ₃ , CAS No. 26628-22-8) which is added as a preservative. Trace amounts of acids and bases have been used only during product processing and to balance the pH of the product. Testing of the product itself to determine health hazards has not been performed. The product does not contain carcinogens.
Labeling:	The product label states the product contains sodium azide (0.09 %) as preservative.
Symbol:	N/A
Hazard Statement:	N/A
Precautionary Statements:	<p>Although sodium azide is only < 0.1 % w/v of the product, the Precautions and Warning section of the product insert states:</p> <ul style="list-style-type: none"> • <i>Harmful if swallowed.</i> • <i>The calibrators / controls contain sodium azide preservative, which may react with lead or copper plumbing to form potentially explosive metal azide. When disposing such liquids always flush with a large volume of water to prevent azide build-up. See Sodium Azide in the National Institute for Occupational Safety (NIOSH). Pocket Guide to Chemical Hazards. Third Printing, September 2007. (https://www.cdc.gov/niosh/npg/default.html)</i> • <i>The calibrators / controls are prepared from non-sterile human urine. They are not tested by licensed reagents for the presence of antibodies to human immunodeficiency viruses, the hepatitis antigens, and/or anti-hepatitis antibodies. They should be handled as potentially infectious. Always apply good laboratory practice to avoid any skin contact or ingestion.</i> • <i>Do not use the calibrators / controls beyond their expiration dates.</i>

3. COMPOSITION / INFORMATION ON INGREDIENTS

There is no hazardous ingredient present in the product which is above its appropriate cutoff levels according to the Globally Harmonized System (GHS). The cutoff level for ingredients causing respiration/skin sensitization, reproductive toxicity, carcinogenicity, and category 1 mutagenicity is $\geq 0.1\%$. The cutoff level for all other hazard classes is $\geq 1\%$. Therefore no components need to be disclosed according to the applicable regulations. The products do contain processed human urine which has not been tested for the presence of antibodies to the human immunodeficiency virus or hepatitis antigens/antibodies. The products should be handled as potentially biohazardous using standard biosafety precautions.

4. FIRST AID MEASURES

Potential Routes of Exposure:	Skin contact, eye contact, and accidental ingestion. Inhalation is highly unlikely.
Skin contact:	Wash affected area thoroughly with soap and water. Remove contaminated clothing and shoes. Flush with plenty of water for at least 15 minutes.
Eye contact:	plenty of water for at least 15 minutes.
Ingestion:	If swallowed, wash out mouth with water and drink plenty of water.

If adverse symptoms are experienced, contact a physician as a precautionary measure.

5. FIRE-FIGHTING MEASURES

Suitable Extinguishing Media:	Typical media is suitable.
Unsuitable extinguishing media:	None are known.
Specific hazards in case of fire:	None are known.
Special protective equipment and precaution for fire fighters:	Use the typical equipment and precautions. There is a recommendation to wear self-contained breathing apparatus.

6. ACCIDENTAL RELEASE MEASURES

Personal Precautions: If the liquid product is spilled, take the proper precautions to minimize exposure to it by wearing the appropriate personal protective equipment (PPE).

Cleaning Methods: Quarantine the spill area and minimize spreading of spilled liquid. Soak up the liquid with a suitable absorbent material such as a mat pad. Clean the spill area thoroughly with soap and water. When disposing of the product by means of plumbing, always flush with a large volume of water to prevent long-term sodium azide accumulation in the metal plumbing. Observe and obey federal, state, and local laws and ordinances regarding proper disposal practices.

Environmental Precautions: Collect the spilled liquid, absorbent material, and rinse water into suitable containers for proper disposal in accordance with applicable waste disposal regulations. Prevent runoff into sewers, storm drains, surface waters, and soil.

7. HANDLING AND STORAGE

Handling: Wear PPE such as gloves, lab coats, and safety glasses as necessary. Avoid contact with eyes or skin.

Storage: Keep the product tightly closed. Store at 2-8 °C. Do not expose to extreme temperatures.

8. EXPOSURE CONTROLS / PERSONAL PROTECTION

Occupational Exposure Limits: There are no established exposure limits for this product.

Personal Protective Equipment (PPE)

Eye Protection: Wear safety glasses with side shields. Have eye wash stations available as a precaution.

Skin Protection: Wear lab coat over clothing and gloves.

Respiratory Protection: Inhalation is unlikely.

9. PHYSICAL AND CHEMICAL PROPERTIES

<i>Physical state:</i>	Calibrators	Controls
<i>Appearance:</i>	Liquid	Liquid
<i>Odor</i>	Transparent	Transparent
<i>pH:</i>	No odor	No odor
<i>Flammable</i>	~6.0 to 7.0	~6.0 to 7.0
<i>Liquid: Solubility:</i>	No	No
<i>Specific Gravity:</i>	Soluble in Water	Soluble in Water
	~1.0	~1.0

10. STABILITY AND REACTIVITY

Chemical Stability: Stable under normal conditions of use.

Possibility of hazardous reactions: Although sodium azide is < 0.1 % w/v of the product, sodium azide build up may occur over time in metal plumbing if the product is not properly disposed, leading to explosive and shock sensitive components. An extra precaution is to schedule a routine assessment of the plumbing system.

Conditions to Avoid: Sodium azide build up in metal plumbing may be avoided by flushing the product with a large volume of water when disposing of it. Additionally, introducing the product to acidic conditions below a pH of 4.8 should be avoided.

Materials to Avoid: None are known.

Hazardous Decomposition Products: None are known.

11. TOXICOLOGICAL INFORMATION

Skin (Irritancy and Acute Toxicity): No data is available.

Eye (Irritancy and Acute Toxicity): No data is available.

Ingestion: No data is available.

Chronic Toxicity (Target Organ Effects): No data is available.

Mutagenicity: No data is available.

Inhalation: No data is available. Low risk of inhalation.

12. ECOLOGICAL INFORMATION

Persistence and Degradability: No data is available.

Bio-accumulative potential: No data is available.

Mobility: No data is available.

Aquatic Toxicity: No data is available.

13. DISPOSAL CONSIDERATIONS

The proper disposal method consists of running a large volume of water with the disposal of the product and/or product wastes in order to prevent sodium azide buildup in metal plumbing. Dispose of the product separately from other waste materials. Follow applicable local, state, and federal regulations for disposal, as some guidelines may require more precautionary measures for disposal of dilute sodium azide liquids, such as segregation of the waste products and treatment.

14. TRANSPORT INFORMATION

These goods are not hazardous and are not regulated for transport by IATA.

Proper Shipping Name: Urinalysis Drugs of Abuse (DAU) Calibrators and Controls

Hazard Class: Non-hazardous but potentially biohazardous

Packing Group: N/A

Marine Pollutant: No

Note: These are perishable, liquid products. The products are packed with ice to maintain an approximately 2-8°C temperature range during transit.

SAFETY DATA SHEET

15. REGULATORY INFORMATION

The regulatory data in this section is not intended to be all-inclusive, only a selected regulation is represented.

European Inventory of Existing Commercial Chemical Substances / European List of Notified Chemical Substances (EINECS / ELINCS) No. for Sodium Azide: 247-852-1.

16. OTHER INFORMATION

SDS Revision: 1

SDS Preparation Date: May 21, 2020

Disclaimer: This information is accurate to the best of MedTest Dx's knowledge. Since the products may be used under conditions beyond the company's control, MedTest Dx does not assume any responsibility for the results of such usage. With the information provided on this SDS, customers receiving MedTest Dx products shall make their own determination of the effects, properties, and measures pertaining to their usage conditions. No warranty or guarantee, expressed or implied, is made concerning the safe use of this product.

Rev. 11

Document number: SDS-POINTEDOACALCTL

Rev. 09/2020